

SPECIFICATION

JUNSEI CHEMICAL CO.,LTD.
QUALITY ASSURANCE SECTION
1-6, Ohmano-cho, Koshigaya-shi,
Saitama, 343-0844, Japan
TEL: +81-48-986-6161
FAX: +81-48-989-2787

Product Code 26160-0310
Commodity Cupferron
Grade Guaranteed Reagent
Q'ty 25g

[Test Item]

Solubility in water
Specific absorbance [$E_{1\text{cm}}^{1\%}$ (281nm)]
Drying loss
Ignition residue (as sulfates)
Sensitivity

[Standard]

CASNo.135-20-6

To pass test
min. 515
max. 1.0 %(m/m)
max. 0.05 %(m/m)
To pass test