

SPECIFICATION

JUNSEI CHEMICAL CO.,LTD.
QUALITY ASSURANCE SECTION
1-6, Ohmano-cho, Koshigaya-shi,
Saitama, 343-0844, Japan
TEL: +81-48-986-6161
FAX: +81-48-989-2787

Product Code 37630-3080
Commodity Distilled water
Grade For HPLC
Q'ty 1L

[Test Item]	[Standard]	CASNo.7732-18-5
Refractive index n_D^{20}	1.332 ~ 1.334	
Non-volatile matter	max. 5 ppm	
Substances reducing $KMnO_4$	To pass test	
Absorbance E 1cm (at 210 ~ 400nm)	max. 0.01	